



Joint-Based Velocity Feedback to Virtual Limb Dynamic Perturbations



Northwestern University

Eric J. Earley^{1,2}, Kyle J. Kaveny², Reva E. Johnson³, Levi J. Hargrove^{1,2}, Jon W. Sensinger⁴

¹Northwestern University, Chicago, IL; ²Shirley Ryan AbilityLab, Chicago, IL; ³Valparaiso University, Valparaiso, IN; ⁴Institute of Biomedical Engineering, University of New Brunswick, Fredericton, NB Canada


Introduction

- Upper-limb amputation reduces sensory feedback, contributing to difficulties performing activities of daily living [1]
- Few attempts to provide complementary sensory feedback have been successful in parallel with vision [2]
- There are three criteria for augmented feedback to be most useful
 - Provide information (i.e. velocity) not available to other senses, notably vision (i.e. position) [3]
 - Feedback should have low uncertainty compared to control of the task
 [4]
 - Provide information in the most uncertain reference frame (i.e. joint-based, rather than egocentric, reference frame) [5]
- These criteria suggest a joint-based velocity feedback paradigm will improve prosthetic arm control, even for those with unaffected vision

Just Noticeable Difference Task


Methods: Two-Alternative Forced Choice

- <u>Motivation</u>: Determine visual discrimination of egocentric angular, joint-based angular, and cartesian speeds
- Two-alternative forced choice with adaptive staircase
 - Subjects shown two sequential movement examples and select example perceived as faster
 - 25 decision reversals, converge on 84% JND



| Tested Object | | Tested Speed | Slow | Medium | Fast | |
|---------------|---------------|--------------------------|--------|--------|--------|--|
| A | Proximal Link | Angular (Egocentric) | 30°/s | 60°/s | 120°/s | |
| В | Distal Link | Angular (Joint-Based) | 30 /8 | 00 / 5 | | |
| С | Endpoint | Cartesian | 2 cm/s | 4 cm/s | 8 cm/s | |


Results




Center-Out Reaching Task


Methods: Two-Arm Linkage with Audio Feedback

- <u>Motivation</u>: Complement egocentric and joint-based speed discrimination with proportional audio feedback
- 2-arm linkage with simulated dynamics
 - Joint torques proportional to enveloped EMG amplitude
 - Audio feedback volume proportional to joint speed



Methods: Center-Out Reaching Task






| | Baseline | No Pert | urbation Perti | urbation L. Ger | neralization R. Gen | eralization | |
|-------------------|----------|---------|----------------|--------------------|---------------------|-------------|--|
| 0 | | 40 | 60 | 80 | 100 | 120 | |
| Original Dynamics | | | | Perturbed Dynamics | | | |

| Testing Block | Baseline | No Perturbation | Perturbation | Left Generalization | Right Generalization |
|---------------|----------|--------------------|--------------|------------------------|-------------------------|
| Trials | 40 | 20 | 20 | 20 | 20 |
| Target | Random | D | D | В | C |

Results



Conclusion

- Discrimination of joint-based speed differences *significantly diminished* when moving slowly compared to egocentric speed
- During center-out reaching task, subjects subjectively reported *improved* awareness of unintentional muscle contracture and increased embodiment of the virtual arm
- Future work includes analysis of JND interaction between egocentric and joint-based linkage speeds, and simplified two-arm linkage control and audio feedback paradigm

Funding

Research supported by NSF-NRI 1317379. E. J. Earley was supported by NIH grant T32 HD07418.

References

- 1. D. S. Childress, "Closed-loop control in prosthetic systems: Historical perspective," *Ann. Biomed. Eng.*, vol. 8, no. 4–6, pp. 293–303, 1980.
- 2. C. Antfolk, M. D'Alonzo, B. Rosén, G. Lundborg, F. Sebelius, and C. Cipriani, "Sensory feedback in upper limb prosthetics.," *Expert Rev. Med. Devices*, vol. 10, no. 1, pp. 45–54, 2013.
- 3. M. M. D. Sobuh, L. P. J. Kenney, A. J. Galpin, S. B. Thies, J. McLaughlin, J. Kulkarni, and P. Kyberd, "Visuomotor behaviours when using a myoelectric prosthesis," *J. Neuroeng. Rehabil.*, vol. 11, no. 1, pp. 1–11, 2014.
- 4. M. O. Ernst and M. S. Banks, "Humans integrate visual and haptic information in a statistically optimal fashion.," *Nature*, vol. 415, no. 6870, pp. 429–433, 2002.
- 5. M. Berniker and K. Kording, "Estimating the sources of motor errors for adaptation and generalization," *Nat. Neurosci.*, vol. 11, no. 12, pp. 1454–1461, 2008.